

Contents

page

Features	2
Quick guide	2
Before first use begins	3
IMPORTANT - Soil pH Pen probe care	3
To operate	5
Information about measuring the pH of soils/media	6
Measuring soil pH value	8
Measuring pH value of a soil solution	9
Cleaning	10
Battery replacement	10
Calibration	11
Error messages	11
Troubleshooting guide	12
Technical specifications	12
Product guarantee	13
Bluelab Probe Care Kits	14
Contact details	14



WEEE-Reg.-Nr. DE 33009570

a handy solution made easy


www.getbluelab.com



Features

Direct pH measurement into soils/media	Low battery warning
Backlit LCD display	Fully waterproof
Hold reading function	Auto off function
Fully guaranteed for 1 year	Selectable units for °C and °F
Integrated storage cap/dibber	Automatic Temperature Compensation (ATC)
Successful pH calibration indicator	

Quick guide



Check mark/tick to indicate successful pH calibration

Disappears 30 days after last successful calibration as a reminder calibration is due.

Low battery warning

Appears when batteries are low.

Power button / hold

Short press to turn on. Short press to hold and unhold reading. Long press to turn off.

Calibrate button

See calibration section.

Units button

Press and hold until units flash then short press to change units. Screen will change back to main display when no buttons have been pressed for 4 seconds.

Storage cap/dibber

The probe must not be allowed to dry out. Always loosen the storage cap/dibber by twisting the dibber before removing or placing back onto probe tip after each use and ensure it contains either BlueLab pH Probe Storage Solution or tap water.

ATTENTION:

FRAGILE Glassware


The pH probe contains a glass tube and a glass bulb.

DO NOT drop, knock or bend.

Refer to section 2.0 - Soil pH Pen probe care.

If it dries, it dies!

Keep your probe tip wet at all times to avoid permanent damage.





1.0 Before first use begins

ATTENTION: Salt-crust forming on the pH probe is normal. A small amount of KCl solution from the storage cap/dibber may leak out leaving a salt-crust. This does not affect the probe and is normal.

The following tasks must be performed before the Bluelab Soil pH Pen is used for the first time.

1 Hydrate Soil pH Pen probe

IMPORTANT: Always loosen the cap before removing or replacing the storage cap/dibber.


- Loosen the Soil pH Pen storage cap/dibber by gripping the top of the storage cap/dibber and gently twisting the bottom of the dibber to loosen slightly. Gently slide the storage cap/dibber off the probe.
- Soak the probe tip in Bluelab pH Storage Solution or fresh tap water for 24 hours. Soaking the probe tip for 24 hours will improve the soil pH probe's activity and is recommended if the probe tip has been allowed to dry.
- Ensure there is always liquid in the storage cap/dibber. Place the loosened storage cap/dibber over the pH probe until seated at pen base, then tighten.

CAUTION: DO NOT hydrate or store the pH probe in Reverse Osmosis (RO), de-ionized or distilled water. When the Soil pH Pen is not in use, place a small amount of Bluelab pH Probe Storage Solution or fresh tap water into the storage cap/dibber.


2 Calibrate the Soil pH Pen before the first use to ensure accurate readings.

Go to section 9.0 for calibration steps.


Removing the Soil pH Pen storage cap/dibber



Hydrating the Bluelab Soil pH Pen



Replacing the Soil pH Pen storage cap/dibber



2.0 IMPORTANT - Soil pH Pen probe care

To ensure you receive a long life from your Soil pH Pen, please ensure you follow the guide below.

The Soil pH Pen probe is fragile

- DO NOT drop, knock or exert a sideways force on the probe.
- Avoid sudden temperature changes.

The soil pH probe does not last forever

- Probes do not last forever. They age through normal use and will eventually fail. The lifetime of a probe depends on the environment it is used in and the way it is treated.
- Always loosen the cap before removing or replacing the storage cap/dibber.
- This probe is suitable to use in temperatures between 0 - 50 °C / 32 - 122 °F only.
- Chemically aggressive, abrasive or oily samples will reduce the life of the probe.

Note: this section continues on next page...



2.0 IMPORTANT - Soil pH Pen probe care cont.

The Soil pH Pen probe requires cleaning at least once a month and prior to calibrating

Cleaning the glassware is critical to obtaining accurate measurements.

Storing the Soil pH Pen

The probe tip must be kept wet - if it dries, it dies!


To prepare the probe for storage, place a small amount of Bluelab pH Probe Storage Solution or tap water (never use RO, distilled or deionized water) into the storage cap/dibber. Place the loosened cap over the probe tip and tighten (see section 1.0).

Long term storage


For long term storage, sit Soil pH Pen upright and ensure the storage cap/dibber contains plenty of Bluelab pH Probe Storage Solution or tap water. Check each month to ensure the liquid has not evaporated.

If the probe has been accidentally allowed to dry out:

The probe must be 'hydrated' for 24 hours in Bluelab pH Probe Storage Solution or fresh clean water (never distilled, RO or deionized water). Following this, carry out a calibration to check if the probe has already suffered permanent damage.



Soil pH Pen body & probe



Storage cap/dibber

DO NOT let the probe tip dry. IF IT DRIES IT DIES!

DO NOT exert sideways force, drop or knock the pen. This will break its external glass bulb or internal glass tube.

DO NOT touch the glass bulb with your fingers as this will contaminate the glass.

DO NOT plunge a cold probe into a hot liquid (or vice versa). Sudden temperature changes can crack the glass and permanently damage the pen.

DO NOT immerse in oils, proteins or suspended solids that will leave a coating on the glass bulb.

DO NOT remove or place the storage cap/dibber back onto the probe without loosening the cap, as you could reduce the life of the probe.



3.0 To operate


1 Turn Soil pH Pen on

Press power button. The last measurement is recalled for 3 seconds.

To turn Soil pH Pen off

Press and hold the power button until OFF is displayed.

NOTE: The Soil pH Pen will automatically turn off after 4 minutes to conserve battery power.




Power button

2 Measure pH

a) With the storage cap/dibber on the pen, press the Soil pH Pen into the sample area until it reaches the word 'soil' at the top of the dibber. Remove gently from soil or substrate, leaving a small hole.


b) Remove storage cap/dibber (refer to section 1.1a for instructions). Turn the Soil pH Pen on, then place pH probe into the newly made hole in the soil or substrate. Do not exert sideways pressure. Wait for the reading to reach a constant value.

NOTE: Never make a hole in the soil or substrate with the pH probe. **Always** use the storage cap/dibber.




3 To hold reading

If you want to "hold" the reading on the screen, short press the power button. To exit the hold function, short press the power button again.




1 second alternating displays



4 To change temperature units


Hold down the units button for 3 seconds until the temperature units start flashing. Release. Short press the units button again to cycle between °F and °C. To exit this mode don't press anything for 3 seconds.

NOTE: You can change units while in hold mode.



5 Rinse and place on the storage cap/dibber

To ensure accurate pH readings always rinse the probe in clean fresh tap water before placing on the storage cap/dibber. Ensure there is BlueLab pH Probe Storage Solution or a little tap water in the storage cap/dibber. Place on the storage cap/dibber, see section 1.1a.





4.0 Information about measuring the pH of soils/media


pH is the measurement of the hydrogen ion concentration (H⁺) - acidity and its opposite, alkalinity. Neutral pH is 7.0 pH. Acidity measures below seven pH (7.0 pH) with alkalinity measuring above it (7.0 pH). See chart below.

In soils or growing media, pH strongly influences the availability of nutrients and the presence of microorganisms in the soil.

Certain plants require a particular pH range to enable the required nutrients to be consistently available to the plant. If the solution is too acidic or too alkaline it can cause “lock up” – a situation which restricts certain elements essential for growth from being absorbed by the root structure. This in turn reduces plant health and performance. Deficiencies in the required elements become apparent in plant growth and can lead to crop failure.

Low soil pH causes aluminium and manganese toxicity in plants and reduces the availability of soil phosphorus. High soil pH also reduces soil phosphorus availability and reduces micro nutrients such as zinc and boron to plants.

The chart below shows how nutrient pH levels influence the uptake of certain elements.





4.0 Information about measuring the pH of soils/media cont.

Taking pH measurements of soils with an electronic meter is indicative rather than absolute.

The following factors are outside the control of any soil pH meter, so to minimise their effect on the accuracy of the pH measurement you should consider the following precautions:

Moisture level/raw water

If the sample you are wishing to measure is dry, add RO water or distilled water to moisten. Ideally wait 24 hours before you take a measurement.

NOTE: If you add tap water, you will influence the soil pH reading based on the pH of the tap water.

Calibration of the Soil pH Pen and cleanliness of the soil probe tip

Calibrating the Soil pH Pen at least monthly will ensure accurate readings. Cleaning the soil residue from the probe tip and storing the Soil pH Pen in a clean moist state will provide reliable readings as well as prolonging the pens life.

Sample selection

For field testing, remove the top 5 - 10 cm / 2 - 4" of the top of the soil. Samples are taken approximately 15 - 20 cm / 6 - 8" down into the substrate and from various areas, then an average of the readings is used.

For container grown plants, it is recommended to check the pH level of the substrate prior to planting.

Factors affecting pH in the soil or media:

Soil type

Soils formed under high rainfall conditions (e.g. Eastern USA) are more acidic than those formed under dry conditions (e.g. Western USA).

Growth stage of the plant

Uptake and requirements of particular elements change as the plant progresses through it's growing cycle. Recording pH level data to create a history is valuable.

Applications and types of fertilizers

Applications and types of fertilizers can alter the pH level significantly. The time at which you take the reading is important. Evaluate the brand of fertilizer to see if it is altering the pH in the wrong direction.

Applications of sprays

As sprays can soak into the soil/media, a change to the pH level could result.

Soil/media temperature

High temperature soils may have a high concentration of CO₂. The higher the concentration of carbon dioxide pressure results in more carbonic acid which lowers pH.

pH range for soil crops

The recommended pH range for soil crops is 6.2 - 7.2, but this is plant specific.



5.0 Measuring soil pH value

Follow the steps below to take pH measurements of soil and media.

1 Remove the top layer of soil/media from the surface of the sample area.

2 Insert the storage cap/dibber into the sample up to the 'soil' indicator as shown on the storage cap/dibber, then remove.

TIP: The dibber creates a safe pathway for the Soil Probe, reducing the likelihood of probe breakage. It should be used at all times.

3 Remove the storage cap/dibber from the pen and turn the Soil pH Pen on.

TIP: to keep the solution or tap water in the storage cap/dibber while taking measurements with the Soil pH Pen, simply place the storage cap/dibber upright into the soil.

4 Insert the probe into the hole made by the dibber, ensuring the probe end makes gentle contact with the soil.




5 Wait for the reading displayed to stabilize to a constant value. Record the reading.

TIP: If the Soil pH Pen turns off while taking a measurement, simply press the power button to turn the Soil pH Pen back on and continue with your measurement.

6 Remove the probe gently from the soil/media and rinse the probe tip under fresh running water (not RO or distilled) to remove any soil residue.

IMPORTANT: Always rinse the probe tip in between each measurement and shake to remove excess water.

7 For field testing, repeat the procedure in different locations and take the average of the measured data as the pH level is representative of the sample area.





6.0 Measuring the pH value of a soil solution

The greatest source of error in soil analysis comes during sample collection. An effort should be made to ensure each sample properly represents the area being sampled.

Collection of sample

- 1 Sample in a zig-zag pattern across the required area.
- 2 Take samples of soil at approximately 20 cm / 8" deep.
- 3 Mix all collected samples together thoroughly.
- 4 Ideally, allow to dry in the air or in an oven at 40 °C / 104 °F.
- 5 Weigh out 20 g / 0.7 oz of the collected soil into a 150 ml / 5 fl oz plastic sample jar.

Sample preparation

- 1 Add 100 ml / 3 fl oz of distilled or deionized water, screw lid on tightly.
- 2 Shake continuously for 5 minutes. Leave overnight and shake again the next morning.
- 3 Allow to settle for 15 minutes after shaking and strain sample into clean measuring cup.

Take pH readings as follows:

- 1 Remove the storage cap/dibber and insert the probe tip into the soil solution sample.
- 2 Turn the Soil pH Pen on.
- 3 Wait for the reading displayed to stabilize to a constant value. Record the reading.


If the Soil pH Pen turns off while taking a measurement, simply press the power button to turn the Soil pH Pen back on and continue with your measurement.
- 4 Remove the probe from the soil solution and rinse the probe tip under fresh running water (not RO or distilled) to remove soil residue.
- 5 Place the storage cap/dibber back onto the probe after use, ensuring it contains BlueLab pH Probe Storage Solution or plenty of tap water.
"IF IT DRIES - IT DIES!". See section 2.0 - Soil pH Probe Care.



7.0 Cleaning

To ensure accurate readings the probe tip needs to be kept clean at all times. Cleaning prior to calibration is necessary for successful calibration.

- 1 **Remove storage cap/dibber.**
Rinse probe tip under fresh tap water.
- 2 **Fill small plastic container with clean water.**
Add a small amount of BlueLab pH Probe Cleaner or mild detergent (dishwashing liquid).
- 3 **Gently stir the probe tip in the mixture.**
Ensure that you do not 'knock' the probe on the side of the container as this may cause damage to the probe. Rinse well under fresh running water to remove all traces of the detergent mixture.
- 4 **If the probe requires removal of heavy contamination:**
Gently brush around the glassware with a few drops of BlueLab pH Probe Cleaner or mild detergent (dishwashing liquid) and a soft toothbrush.
- 5 **Rinse well under fresh running water to remove all traces of the detergent mixture.**
- 6 **Calibration of the probe is required after every clean.**
See the pH calibration in section 9.0. Place storage cap/dibber back onto the probe tip.



8.0 Battery replacement

The pH pen is powered with 1 x AAA alkaline battery. Do not use rechargeable batteries. A low battery warning is indicated by a battery symbol appearing on the screen. Only remove the battery cap when the batteries require changing. Battery life is expected to be approx. 350 hours.

- 1 **To remove old battery**
Undo battery cap fasteners. Remove battery cap and tip out the old battery.
- 2 **Fit new battery**
Insert the new battery positive (+) end down into the body.
- 3 **Ensure waterproof battery cap seal is clean, and free from any debris.**
Seal will fail if any dirt or debris is present.
- 4 **Replace battery cap**
Tighten fasteners on battery cap. Do not over tighten.
Ensure silicone seal on battery cap is fully enclosed in pen body. This ensures the unit remains 100% waterproof.





9.0 Calibration

pH calibration is required before first use to ensure that the first reading is accurate. Calibration is also required when:

- The check mark/tick has disappeared from the LCD screen (30 days after last successful calibration)
- The reading is different from what you expected
- After cleaning
- After changing the battery

pH 7.0 and pH 4.0 solutions are required for calibration and should be decanted into small, clean plastic containers. You may also calibrate using pH 7.0 and pH 10.0 solutions if your readings will normally be higher than 7.0 pH.

1 Excluding first use, YOU MUST CLEAN the probe tip before calibrating.

See section 7.0 Cleaning.

2 Remove storage cap/dipper.

Refer to section 1.0, step 1.

3 Rinse probe tip in fresh water and place in pH 7.0 solution.

Wait for reading to stabilize to a constant value.

4 Press the cal button until CAL is displayed.

Release button. When CAL 7 is displayed, 1 point calibration is achieved.


5 Rinse probe in fresh water and place it in either pH 4.0 or pH 10.0 solution (use pH 10.0 solution if you expect to measure above 7.0 pH).

Wait for reading to stabilize to a constant value.

6 Press the cal button until CAL 4 or CAL 10 is displayed.


CAL 4 or CAL 10 should be displayed (depending on what solution you are calibrating in). The check mark/tick is displayed when a 2 point (or 3) calibration is completed.

NOTE: For a three point calibration repeat steps 2, 3 & 4 using pH 7.0, 4.0 then 10.0 solution.



10.0 Error messages

The following error messages appear for the following reasons.



Temperature under range



pH over range




Temperature over range



pH under range



pH calibration failed



Hardware error



11.0 Troubleshooting guide

Trouble	Reason	Correction
<i>Drift - readings slowly varying</i>	Glassware not clean	Clean glassware and calibrate
	Wick contaminated, blocked or dry	Soak probe in BlueLab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required
	Glassware aged	Replace unit
<i>Displays similar pH reading in all buffers no matter what the buffer value is</i>	Glassware cracked or broken	Replace unit
<i>Unsuccessful pH calibration</i>	Buffers inaccurate	Replace buffers
	Glassware not clean	Clean glassware
	Glassware aged (glassware will not clean)	Replace unit
	Probe not hydrated	Soak probe in BlueLab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required
<i>Noisy - readings jumping</i>	Contact zone not immersed	Ensure soil/substrate is damp
	Probe not adequately hydrated	Soak probe in BlueLab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required
<i>Displays pH 7 for all buffers</i>	Glassware broken	Replace unit
<i>Incorrect sample reading following successful pH calibration</i>	Wick blocked	Soak probe in BlueLab pH Probe Storage Solution or tap water only for 24 hours and retest. Do not measure proteins or oils with this unit. Replacement of unit may be required

12.0 Technical specifications

Measurement range	0.0 - 14.0 pH, 0 - 50 °C, 32 - 122 °F
Resolution	0.1 pH, 1 °C / 1 °F
Accuracy at 25 °C / 77 °F	± 0.1 pH, ± 1 °C / ± 2 °F
Temperature compensation	Automatic
Operating temperature	0 - 50 °C, 32 - 122 °F
Calibration	Two point, pH 7.0 and pH 4.0 or pH 10.0
Units	pH, °F and °C
Power source	1 x AAA alkaline battery



Bluelab Soil pH Pen product guarantee

Bluelab Corporation Limited guarantees this product for a period of **1 year (12 months)** from the date of sale to the original purchaser. The product will be repaired or replaced, should it be found faulty due to component failure, or faulty workmanship. The faulty product should be returned to the point of purchase.



The guarantee is null and void should any internal parts or fixed external parts be tampered with or altered in any way, or should the unit have been incorrectly operated, or in any way be maltreated. This guarantee does not cover reported faults which are shown to be caused by any or all of the following: contaminated measuring tip (see instruction manual for cleaning instructions), broken glassware or drying of the pH probe glassware, flat or damaged batteries or batteries that have been incorrectly inserted, or damaged battery contacts or connections caused by incorrect battery replacement or ingress of moisture from incorrect positioning of the battery cap and waterproof seal.

NO RESPONSIBILITY will be accepted by Bluelab or any of its agents or resellers should any damage or unfavourable conditions result from the use of this product, should it be faulty or incorrectly operated.

Register your guarantee online at: www.getbluelab.com

Or fill out the form below and post, email or fax to:

Bluelab Corporation Limited 8 Whiore Ave, Tauriko Industrial Park, Tauranga 3110, New Zealand
Fax: +64 7 578 0847 **Email:** support@getbluelab.com

Product details

Product name	
Serial number	
Date purchased	

Purchaser details

Purchaser's name	
Address	
City	
Country	
Email (optional)	

Purchased from (Dealers details)

Purchased from	
Address	
City	
Country	
Phone number (optional)	

Limitation of Liability

Under no circumstances shall Bluelab Corporation Limited be liable for any claims, losses, costs and damages of any nature whatsoever (including any consequential loss) that result from the use of, or the inability to use, these instructions.



Bluelab Probe Care Kits

The instrument is only as accurate as the probe is clean!

Probe cleaning is one of the most important parts of owning and operating any Bluelab meter, monitor or controller. If the probe is contaminated (dirty) it affects the accuracy of the reading displayed.

The probe surface is where the instrument takes the reading of the solution. The information is sent back from the probe to the electronic brain of the instrument. A calculation is then done in the instrument's brain or micro computer and a reading is displayed. If the information sent back from the probe is inaccurate due to probe surface contamination then the reading will be inaccurate. Cleaning the probes is a very easy task and will prolong the life of the probes.



Bluelab Probe Care Kit - pH contents:



- › Cleaning instructions inside box lid
- › 500ml pH4 and pH7 Calibration solutions
- › Decanter vessels
- › Bluelab pH Probe Cleaner
- › Toothbrush (probe cleaning instrument)

Bluelab Probe Care Kit - Conductivity contents:



- › Cleaning instructions inside box lid
- › 500ml 2.77EC conductivity standard solution
- › Decanter vessel
- › Bluelab Conductivity Probe Cleaner
- › Bluelab Chamois (probe cleaning instrument)



If you need assistance or advice - we're here to help you.
Phone: **+64 7 578 0849** Fax: **+64 7 578 0847**
Email: **support@getbluelab.com**



Looking for specifications or technical advice?
Visit us online @ **www.getbluelab.com**



Bluelab Corporation Limited
8 Whiore Avenue, Tauriko Industrial Park
Tauranga 3110, New Zealand



Instruction Manual English PENSOLPH_V01_071112
© Copyright 2012, all rights reserved, Bluelab Corporation Limited